
BACHARACH 2011 PRICE LIST

INSTRUMENT CALIBRATION SERVICE

<u>Model</u>	<u>Calibration Service</u>	<u>Calibration Price</u>
Leakator Jr	Functional Test	\$110.00
Leakator 10	Functional Test	\$110.00
Monoxor II	CO	\$110.00
Monoxor III	CO	\$110.00
Fyrite Tech 50	O2, Thermocouples	\$165.00
Fyrite Tech 60	CO, O2, Thermocouples	\$165.00
Fyrite Pro 105	CO, Draft	\$165.00
Fyrite Pro 110	O2, Thermocouples	\$165.00
Fyrite Pro 120	CO, O2, Thermocouples	\$165.00
Fyrite Pro 125	CO, O2, Draft, Thermocouples	\$165.00
Fyrite Insight	CO, O2, Draft, Thermocouples	\$165.00
PCA 20	CO, Thermocouples	\$120.00
PCA 25 & 55	CO, Draft, Thermocouples	\$120.00
PCA 35 & 65	NO, CO, Draft, Thermocouples	\$150.00
PCA-2 225	CO, Draft, Thermocouples	\$220.00
PCA-2 235	NO, CO, Draft, Thermocouples	\$220.00
PCA-2 245	CO-High, CO, Draft, Thermocouples	\$220.00
PCA-2 255	SO2, CO, Draft, Thermocouples	\$220.00
PCA-2 265	NO2, NO, CO, Draft, Thermocouples	\$220.00
PCA-2 275	SO2, NO, CO, Draft, Thermocouples	\$220.00
PCA-2 285	NO, CO-High, CO, Draft, Thermocouples	\$220.00
ECA 450 Base Unit	CO, Draft, Thermocouples	\$220.00
ECA 450 Each Added Sensor		\$55.00

NOTE: Refer to your instrument manual for recommended calibration intervals. Your instrument should be calibrated at least every six months, if not more frequently, to maintain the accuracy listed in Bacharach's Technical Characteristics Section of the instrument's manual. The Fyrite Pro and Fyrite Tech O2 sensors are calibrated to 100% N2 and 20.9% O2 in ambient air. The Fyrite Insight, PCA, PCA-2, and ECA 450 O2 sensors are self-calibrating. All standard CO sensors are calibrated to 500ppm CO; CO-High to 4000ppm. The PCA, PCA-2, and ECA standard CO sensor is also calibrated to 1000ppm CO and 1000ppm H2 in Nitrogen. A Certificate of Calibration is included. Calibration prices do not include the material and labor cost of any needed sensors or parts, or shipping and handling to return the instrument to you. Please e-mail service@entechsupply.com to receive a Bacharach Calibration and Non-Warranty Service Request Form.